

IMR Test Lab

Chemical Resistance Test

Arpa USA,
13300 Vantage Way
Jacksonville, FL., 32218

IMR Report
#200600783 for SEFA-
8 Procedure



Test Number	Chemical Reagent	Rating Level/Labolam	Rating Level Compact Grade
1	Acetate, Amyl	0	0
2	Acetate, Ethyl	0	0
3	Acetic Acid, 98%	0	0
4	Acetone	0	1
5	Acid Dichromate, 5%	0	0
6	Alcohol, Butyl	0	0
7	Alcohol, Ethyl	0	0
8	Alcohol, Methyl	0	0
9	Ammonium Hydroxide, 28%	0	0
10	Benzene	0	1
11	Carbon Tetrachloride	0	0
12	Chloroform	0	0
13	Chromic Acid, 60%	0	0
14	Cresol	0	0
15	Dichlor Acetic Acid	0	0
16	Dimethylformamide	0	0
17	Dioxane	0	0
18	Ethyl Ether	1	1
19	Formaldehyde, 37%	0	0
20	Formic Acid, 90%	1	1
21	Furfural	0	0
22	Gasoline	0	0
23	Hydrochloric Acid, 37%	1	0
24	Hydrofluoric Acid, 48%	1	1
25	Hydrogen Peroxide, 3%	0	0
26	Iodine, Tincture of	0	0
27	Methyl Ethyl Ketone	0	0
28	Methylene Chloride	0	0
29	Mono Chlorobenzene	0	0
30	Napthalene	0	0
31	Nitric Acid, 20%	1	0
32	Nitric Acid, 30%	1	0
33	Nitric Acid, 70%	2	1
34	Phenol, 90%	1	0
35	Phosphoric Acid, 85%	0	0
36	Silver Nitrate, Saturated	2	2

37	Sodium Hydroxide, 10%	0	0
38	Sodium Hydroxide, 20%	0	0
39	Sodium Hydroxide, 40%	0	0
40	Sodium Hydroxide, Flake	0	0
41	Sodium Sulfide, Saturated	0	0
42	Sulfuric Acid, 33%	0	0
43	Sulfuric Acid, 77%	1	1
44	Sulfuric Acid, 96%	2	2
45	Sulfuric Acid (77%) and Nitric Acid (70%), Equal Parts	3	3
46	Toluene	0	1
47	Trichloroethylene	0	0
48	Xylene	0	0
49	Zinc Chloride, Saturated	0	0